

TCB

**GRANT OF EQUIPMENT
AUTHORIZATION**

TCB

Certification

**Issued Under the Authority of the
Federal Communications Commission**

By:

**EMCCert Dr. Rasek GmbH
Stoernhofer Berg 15
91364 Unterleinleiter,
Germany**

Date of Grant: 04/29/2016

Application Dated: 04/29/2016

**Carstens Medizinelektronik GmbH
Nelkenweg 8
Bovenden - Lower Saxony, 37120
Germany**

Attention: Bahne Carstens

NOT TRANSFERABLE

EQUIPMENT AUTHORIZATION is hereby issued to the named GRANTEE, and is VALID ONLY for the equipment identified hereon for use under the Commission's Rules and Regulations listed below.

FCC IDENTIFIER: 2AHFP-AG501
Name of Grantee: Carstens Medizinelektronik GmbH
Equipment Class: Part 15 Low Power Communication Device
Transmitter
Notes: 3D Electromagnetic Articulograph

<u>Grant Notes</u>	<u>FCC Rule Parts</u>	<u>Frequency Range (MHZ)</u>	<u>Output Watts</u>	<u>Frequency Tolerance</u>	<u>Emission Designator</u>
19	15C	0.0075 - 0.01375			

Two ferrites (Würth 742 711 31) at the endings of the Ethernet must be attached to the interface cable.

19: This device must be marketed with a shielded interface cable which incorporates ferrite cores equal in quantity and type to those used during Certification testing.

